

## IR Spectra of Carbon paper (right and wrong side)

Both sides of the carbon paper are measured by ATR method. Completely different spectra are obtained.

### Measuring conditions

Measurement method : ATR methodmethod  
ATR prism : Ge prism  
Number of scans : 64 times  
Resolution :  $4\text{ cm}^{-1}$   
Detector : TGS detector

