

IR Spectra of Thermal papers (new and old)

New and old thermal papers are cut in the size of ATR prism and measured. And a differential spectrum is calculated.

Measuring conditions

Measurement method : ATR method
ATR prism : KRS-5 prism
Number of scans : 64 times
Resolution : 4 cm^{-1}
Detector : TGS detector

